

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/14/2019

OPERATOR NAME:
CORBET W SULLIVAN
Permit #: 45898
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:21
INTERNAL STANDARD	VERIFIED	09:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:22
BLANK TEST	0.00	09:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:23
BLANK TEST	0.00	09:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:25
BLANK TEST	0.00	09:26

CS #623

