

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/14/2019

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:42
INTERNAL STANDARD	VERIFIED	15:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:42
BLANK TEST	0.00	15:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:44
BLANK TEST	0.00	15:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:45
BLANK TEST	0.00	15:46

Q #691

