

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/02/2019

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:55
INTERNAL STANDARD	VERIFIED	22:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:56
BLANK TEST	0.00	22:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:57
BLANK TEST	0.00	22:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:58
BLANK TEST	0.00	22:59

Q #691

