

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/07/2019

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:25
INTERNAL STANDARD	VERIFIED	17:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:25
BLANK TEST	0.00	17:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:26
BLANK TEST	0.00	17:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:28
BLANK TEST	0.00	17:29

CM #821

