

SUPERVISOR TEST

WASHINGTON COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106406

Date: 12/09/2019

OPERATOR NAME:  
 KYLE L KOLAR  
 Permit #: 44063  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:51
INTERNAL STANDARD	VERIFIED	23:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:51
BLANK TEST	0.00	23:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:53
BLANK TEST	0.00	23:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:54
BLANK TEST	0.00	23:55

*#691*

