

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 02/11/2019

OPERATOR NAME:  
TRAVIS A EARL  
Permit #: 33337  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:47
INTERNAL STANDARD	VERIFIED	06:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:48
BLANK TEST	0.00	06:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:52

*Travis #650*

