

SOLUTION CHANGE

-----  
 WASHINGTON COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106406  
 -----

Date: 03/02/2019

OPERATOR NAME:  
 JOHN K LATOUR  
 Permit #: 31414  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:01
INTERNAL STANDARD	VERIFIED	00:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:03
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:06

#653  
 JKL

