

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 04/08/2019

OPERATOR NAME:  
KATHERINE M PADDOCK  
Permit #: 45899  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:26
INTERNAL STANDARD	VERIFIED	14:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:26
BLANK TEST	0.00	14:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:28
BLANK TEST	0.00	14:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:29
BLANK TEST	0.00	14:30

*KR*  
*#1013*

