

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 05/01/2019

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:55
INTERNAL STANDARD	VERIFIED	15:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:56
BLANK TEST	0.00	15:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:57
BLANK TEST	0.00	15:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:58
BLANK TEST	0.00	15:59

CM #821

