

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 05/27/2019

OPERATOR NAME:
WILLIAM L HOLCOMB
Permit #: 31942
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 1136 19030 QIA

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:16
INTERNAL STANDARD	VERIFIED	22:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:16
BLANK TEST	0.00	22:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:18
BLANK TEST	0.00	22:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:19
BLANK TEST	0.00	22:20

Sgt W. Holcomb #629

