

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 06/10/2019

OPERATOR NAME:  
KYLE L KOLAR  
Permit #: 44063  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:08
INTERNAL STANDARD	VERIFIED	21:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:08
BLANK TEST	0.00	21:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:10
BLANK TEST	0.00	21:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:11
BLANK TEST	0.00	21:12

*Q # 691*

