

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/01/2019

OPERATOR NAME:
JENNIFER M NOBLE
Permit #: 44064
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:00
INTERNAL STANDARD	VERIFIED	17:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:01
BLANK TEST	0.00	17:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:02
BLANK TEST	0.00	17:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:03
BLANK TEST	0.00	17:04

Jennifer Noble

