

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/26/2019

OPERATOR NAME:
CORY L HURLBURT
Permit #: 37017
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:43
INTERNAL STANDARD	VERIFIED	13:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:44
BLANK TEST	0.00	13:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:45
BLANK TEST	0.00	13:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:47
BLANK TEST	0.00	13:48

CH # 757

