

SUPERVISOR TEST

WASHINGTON COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106406

Date: 09/30/2019

OPERATOR NAME:  
 CHRISTOPHER J MURRAY  
 Permit #: 46226  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 19130

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 08:28 |
| INTERNAL STANDARD       | VERIFIED | 08:28 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.09     | 08:28 |
| BLANK TEST              | 0.00     | 08:29 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.09     | 08:30 |
| BLANK TEST              | 0.00     | 08:31 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.09     | 08:31 |
| BLANK TEST              | 0.00     | 08:32 |

*Christopher J Murray*  
 #829

