

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 01/30/2019

OPERATOR NAME:  
LANCE M BARBER  
Permit #: 42841  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:11
INTERNAL STANDARD	VERIFIED	04:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:12
BLANK TEST	0.00	04:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:13
BLANK TEST	0.00	04:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:14
BLANK TEST	0.00	04:15

*Handwritten signature and initials*

