

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 10/21/2019

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:26
INTERNAL STANDARD	VERIFIED	23:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:27
BLANK TEST	0.00	23:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:28
BLANK TEST	0.00	23:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:29
BLANK TEST	0.00	23:30

