

SOLUTION CHANGE

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 11/01/2019

OPERATOR NAME:  
LANCE M BARBER  
Permit #: 42841  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:48
INTERNAL STANDARD	VERIFIED	06:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:50
BLANK TEST	0.00	06:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:52

