

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/05/2019

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:55
INTERNAL STANDARD	VERIFIED	10:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:56
BLANK TEST	0.00	10:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:57
BLANK TEST	0.00	10:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:58
BLANK TEST	0.00	10:59

