

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 03/13/2019

OPERATOR NAME:
JASON R FLYNN
Permit #: 45901
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:10
INTERNAL STANDARD	VERIFIED	16:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:11
BLANK TEST	0.00	16:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:12
BLANK TEST	0.00	16:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:13
BLANK TEST	0.00	16:14

J. R. Flynn
007

