

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 03/05/2019

OPERATOR NAME:
JOSHUA T WHITNEY
Permit #: 40382
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:23
INTERNAL STANDARD	VERIFIED	08:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:24
BLANK TEST	0.00	08:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:25
BLANK TEST	0.00	08:26

WHITNEY 695

