

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 07/22/2019

OPERATOR NAME:
MICHAEL O WEBER
Permit #: 46067
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:31
INTERNAL STANDARD	VERIFIED	15:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:32
BLANK TEST	0.00	15:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:33
BLANK TEST	0.00	15:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:34
BLANK TEST	0.00	15:35

[Handwritten Signature]
6/11

