



**State
Police**

**Forensic
Investigation
Center**

ANDREW M. CUOMO

Governor

KEITH M. CORLETT

Superintendent

DR. RAY A. WICKENHEISER

Director

0.10% SIMULATOR SOLUTION RECORD

Report Date: September 3, 2019
Manufacturer: Guth Laboratories, Inc.
Lot Number: 19180
Expiration Date: 1/24/2020
Analysis Results: 122 mg Ethyl Alcohol / 100 mL of solution

I, Carrie A Kirkton, tested simulator solution lot number 19180 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carrie A Kirkton, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

 Digitally signed under
 ESRA by Carrie A Kirkton on
 2019.09.03 10:40:09 -04'00'

Carrie A Kirkton
Supervisor of Forensic Services
Breath Analysis Technical Supervisor

Reviewed by:

 Digitally signed under
 ESRA by Jennifer F Limoges
 on
 2019.09.03 14:26:34 -04'00'

Jennifer F Limoges
Associate Director of Forensic Science
Breath Analysis Technical Supervisor