

SUPERVISOR TEST

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 12/11/2019

OPERATOR NAME:
 RICHARD J ARNOLD
 Permit #: 33860
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:40
INTERNAL STANDARD	VERIFIED	18:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:40
BLANK TEST	0.00	18:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:41
BLANK TEST	0.00	18:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:43
BLANK TEST	0.00	18:44

