

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 07/10/2019

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:20
INTERNAL STANDARD	VERIFIED	20:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:20
BLANK TEST	0.00	20:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:21
BLANK TEST	0.00	20:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:23
BLANK TEST	0.00	20:23

