

SUPERVISOR TEST

-----  
HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408  
-----

.te: 07/24/2019

OPERATOR NAME:  
BRAD M LYON  
Permit #: 44978  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:44
INTERNAL STANDARD	VERIFIED	02:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:44
BLANK TEST	0.00	02:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:46
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:47
BLANK TEST	0.00	02:48

