

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 07/31/2019

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:14
INTERNAL STANDARD	VERIFIED	00:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:15
BLANK TEST	0.00	00:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:16
BLANK TEST	0.00	00:17
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.10	00:17
BLANK TEST	0.00	00:18

