

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 08/12/2019

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:48
INTERNAL STANDARD	VERIFIED	19:48
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.10	19:48
BLANK TEST	0.00	19:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:49
BLANK TEST	0.00	19:50
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.10	19:51
BLANK TEST	0.00	19:52

