

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 09/04/2019

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:33
INTERNAL STANDARD	VERIFIED	19:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:34
BLANK TEST	0.00	19:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:35
BLANK TEST	0.00	19:36
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.10	19:36
BLANK TEST	0.00	19:37

