

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 10/23/2019

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:45
INTERNAL STANDARD	VERIFIED	06:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:46
BLANK TEST	0.00	06:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:47
BLANK TEST	0.00	06:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:48
BLANK TEST	0.00	06:49

