

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/30/2019

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:00
INTERNAL STANDARD	VERIFIED	01:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:01
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:02
BLANK TEST	0.00	01:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:03
BLANK TEST	0.00	01:04

