

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 01/06/2020

OPERATOR NAME:  
MICHAEL O WEBER  
Permit #: 46067  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:57
INTERNAL STANDARD	VERIFIED	18:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:58
BLANK TEST	0.00	18:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:59
BLANK TEST	0.00	19:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:00
BLANK TEST	0.00	19:01

