

SOLUTION CHANGE

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 08/01/2019

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:07
INTERNAL STANDARD	VERIFIED	15:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:07
BLANK TEST	0.00	15:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:09
BLANK TEST	0.00	15:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:10
BLANK TEST	0.00	15:10

MKJ #643

