

SOLUTION CHANGE




WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 02/03/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:22
INTERNAL STANDARD	VERIFIED	07:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:25
BLANK TEST	0.00	07:25



SUPERVISOR TEST


WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 01/25/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:33
INTERNAL STANDARD	VERIFIED	06:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:33
BLANK TEST	0.00	06:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:34
BLANK TEST	0.00	06:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:36
BLANK TEST	0.00	06:36



SUPERVISOR TEST




WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 1.13806

Date: 01/18/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:52
INTERNAL STANDARD	VERIFIED	13:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:53
BLANK TEST	0.00	13:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:54
BLANK TEST	0.00	13:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:55
BLANK TEST	0.00	13:56



SUPERVISOR TEST



WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 01/11/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:02
INTERNAL STANDARD	VERIFIED	22:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06



SOLUTION CHANGE

-----  
 WHITEHALL VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 113806  
 -----

Date: 01/04/2020

OPERATOR NAME:  
 BRYAN J GRECO  
 Permit #: 44077  
 Operator Agency/Dept: WHITEHALL VILL. P.D.  
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:30
INTERNAL STANDARD	VERIFIED	14:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:30
BLANK TEST	0.00	14:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:31
BLANK TEST	0.00	14:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:33
BLANK TEST	0.00	14:33

