

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 01/28/2020

OPERATOR NAME:  
CORY L HURLBURT  
Permit #: 37017  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:42
INTERNAL STANDARD	VERIFIED	06:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:43
BLANK TEST	0.00	06:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:44
BLANK TEST	0.00	06:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:45
BLANK TEST	0.00	06:46

*Handwritten:* #257

