

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 02/05/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:04
INTERNAL STANDARD	VERIFIED	19:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:04
BLANK TEST	0.00	19:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:06
BLANK TEST	0.00	19:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:07
BLANK TEST	0.00	19:08

