

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 02/08/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:25
INTERNAL STANDARD	VERIFIED	22:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:26
BLANK TEST	0.00	22:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:27
BLANK TEST	0.00	22:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:28
BLANK TEST	0.00	22:29

