

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 02/10/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:18
INTERNAL STANDARD	VERIFIED	07:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:19
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:21
BLANK TEST	dmt: 126706	0.00 07:22

Date: 02/10/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:18
INTERNAL STANDARD	VERIFIED	07:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:19
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:21
BLANK TEST	0.00	07:22

