

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 01/27/2020

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:18
INTERNAL STANDARD	VERIFIED	20:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:19
BLANK TEST	0.00	20:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:20
BLANK TEST	0.00	20:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:21
BLANK TEST	0.00	20:22

