

SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 02/03/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:50
INTERNAL STANDARD	VERIFIED	20:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:51
BLANK TEST	0.00	20:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:52
BLANK TEST	0.00	20:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:53
BLANK TEST	0.00	20:54

