

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 02/10/2020

OPERATOR NAME:  
CORBET W SULLIVAN  
Permit #: 45898  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:57
INTERNAL STANDARD	VERIFIED	06:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:58
BLANK TEST	0.00	06:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:59
BLANK TEST	0.00	07:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:00
BLANK TEST	0.00	07:01

CS #623

