

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 02/17/2020

OPERATOR NAME:  
KYLE L KOLAR  
Permit #: 44063  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:56
INTERNAL STANDARD	VERIFIED	16:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:56
BLANK TEST	0.00	16:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:57
BLANK TEST	0.00	16:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:59
BLANK TEST	0.00	17:00

*[Handwritten signature]*  
#691

