

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 02/24/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:19
INTERNAL STANDARD	VERIFIED	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:19
BLANK TEST	0.00	07:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:22
BLANK TEST dmt: 126706	0.00	07:22

Date: 02/24/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:19
INTERNAL STANDARD	VERIFIED	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:19
BLANK TEST	0.00	07:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:22
BLANK TEST	0.00	07:22

