

SOLUTION CHANGE

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 03/07/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:07
INTERNAL STANDARD	VERIFIED	15:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:07
BLANK TEST	0.00	15:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:08
BLANK TEST	0.00	15:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:09
BLANK TEST	0.00	15:10

