

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATA MASTER dmt: 126706

Date: 03/09/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19350

Normal = 0.10 g/210L

BLANK TEST	0.00	07:03
INTERNAL STANDARD	VERIFIED	07:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:04
BLANK TEST	0.00	07:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:05
BLANK TEST	0.00	07:06

