

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DALLMASTER dmt: 126706

Date: 03/16/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:37
INTERNAL STANDARD	VERIFIED	07:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:38
BLANK TEST	0.00	07:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:39
BLANK TEST	0.00	07:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:40
BLANK TEST	0.00	07:41

