

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 03/21/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:58
INTERNAL STANDARD	VERIFIED	10:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:58
BLANK TEST	0.00	10:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:59
BLANK TEST	0.00	11:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:00
BLANK TEST	0.00	11:01

