

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 03/07/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:55
INTERNAL STANDARD	VERIFIED	19:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:55
BLANK TEST	0.00	19:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:56
BLANK TEST	0.00	19:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:57
BLANK TEST	0.00	19:58

