

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

ate: 03/14/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:13
INTERNAL STANDARD	VERIFIED	19:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:14
BLANK TEST	0.00	19:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:15
BLANK TEST	0.00	19:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:16
BLANK TEST	0.00	19:17

