

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 03/25/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:31
INTERNAL STANDARD	VERIFIED	00:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:32
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:34
BLANK TEST	0.00	00:35

