

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 03/24/2020

OPERATOR NAME:
DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:37
INTERNAL STANDARD	VERIFIED	13:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:38
BLANK TEST	0.00	13:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:39
BLANK TEST	0.00	13:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:40
BLANK TEST	0.00	13:41

